



# Test report No 111-18954

<i>Object</i>	<b>Goniometer G-01L</b> Manufacturing date 10.2019 Serial No. 191001
<i>Order</i>	Performance test according to procedure as specified by the applicant
<i>Applicant</i>	<b>ZG Optique SA</b> Fin-de-Praz 24 CH-2024 St. Aubin
<i>Traceability</i>	The reported measurement values are traceable to national standards and thus to internationally supported realizations of the SI-units.
<i>Date of test measurements</i>	27 November 2019
<i>Marking</i>	Calibration label METAS 11.2019
CH-3003 Bern-Wabern, 4 December 2019	
<i>For the Measurements</i>	Dr Rudolf Thalmann
<i>Approved by</i>	Dr Felix Meli, Head of Laboratory Laboratory for length, nano- und microtechnology

## Test Report No 111-18954

### Test procedure and conditions

The measurements and the evaluation of the data have been carried out according to a procedure specified by the applicant (see appendix to this report). The optical polygon used for these measurements was 8-sided, from glass, and calibrated at METAS at the same time using independent equipment (certificate no 111-18953). The entire procedure was performed two times. In the following, the results of the second measurement are reported.

The laboratory temperature during the measurements was  $(20 \pm 0.5)$  °C.

### Measurement results

#### *Average standard deviation*

$$u_1 = 0.024''$$

#### *Type B uncertainty contributions*

$$u_2 = 0.06''$$

$$u_3 = 0.035''$$

#### *Combined standard uncertainty*

$$u_c = 0.07''$$

#### *Comparison of polygon measurements*

Pitch angles between adjacent faces

Position	$\tilde{\varphi}_i$ (G-01L)	$\varphi_i^{tr}$ (METAS)	$\delta = \tilde{\varphi}_i - \varphi_i^{tr}$
1-2	-0.13	$-(0.10 \pm 0.10)''$	-0.04''
2-3	0.81	$(0.83 \pm 0.10)''$	-0.01''
3-4	1.88	$(1.82 \pm 0.10)''$	0.06''
4-5	-0.81	$-(0.76 \pm 0.10)''$	-0.05''
5-6	-1.97	$-(1.99 \pm 0.10)''$	0.01''
6-7	0.35	$(0.35 \pm 0.10)''$	0.00''
7-8	-0.27	$-(0.29 \pm 0.10)''$	0.01''
8-1	0.10	$(0.13 \pm 0.10)''$	-0.03''

## Appendix to Test Report No 111-18954

### Test procedure for Automatic Goniometer GA-1L in goniometer mode

1. The procedure of the performance test should be carried out by means of standard n-faced optical polygon (OP) calibrated by the Federal Institute of Metrology (METAS) or by the Physikalisch-Technische Bundesanstalt (PTB).
2. The procedure of performance test considers the evaluation of the following parameters:
  - Standard uncertainty;
  - Combined standard uncertainty.
3. The central angles  $\varphi_i$  of the OP are measured.
4. The OP is turned relative to the rotary table by the nominal angle between two adjacent faces of the OP.
5. The central angles of the OP are measured again.
6. Then the points 4 and 5 are repeated until  $n$  measurement are made, where  $n$  is the number of the OP faces. As a result, we get an array of angle values  $\varphi_{im}$  where  $i$  is a number of angle between faces  $i$  and  $i + 1$ , and  $m$  is the number of the OP position.
7. The mean values of the angle  $\tilde{\varphi}_i$ s are estimated using the averaging upon the different positions of the OP:

$$\tilde{\varphi}_i = \frac{1}{n} \sum_{m=1}^n \varphi_{im}$$

8. The uncertainty (type A) is determined as average of standard deviations  $\sigma_{im}$  over all faces and all positions

$$u_1 = \sqrt{\frac{\sum_i^n \sum_m^n \sigma_{im}^2}{n \cdot n}}$$

9. The uncertainty (type B) is determined as

$$u_2 = \frac{\max |(\tilde{\varphi}_i - \varphi_{im})|}{\sqrt{3}},$$

$$u_3 = \frac{\max |(\tilde{\varphi}_i - \varphi_i^{tr})|}{\sqrt{3}}$$

where  $\varphi_i^{tr}$  are the METAS calibrated values of central angles of the OP.

10. The combined standard uncertainty is determined as

$$u_c = \sqrt{u_1^2 + u_2^2 + u_3^2}$$

Note: The single measurement for G-01L is considered as the measurement during 16 revolutions of the rotary table.  $\varphi_{im}$  and  $\sigma_{im}$  are the results of the single measurement of the central angle value and the standard deviation, correspondingly.

Prof. Dr. Yuri Filatov